

OFFICE OF THE PROJECT DIRECTOR,
IMPLEMENTATION OF MASTER PLAN
OF SAFARI ZOO LAHORE,
PUNJAB WILDLIFE AND PARKS
DEPARTMENT, LAHORE.

TENDER DOCUMENT

For

PROVISION OF MEDICAL EQUIPMENTS

Date of BID Submission

Component A (Lot Wise)

04-01-2024

Component B (Lot Wise)

05-01-2024

- *Invitation to Bid*
- *Instructions to Bidders*
- *Technical Specifications*
- *Bid Forms*
- *Conditions of Contract*

The bidder is expected to examine the Bidding Documents, including all instructions, forms, terms, specifications and charts/drawings. Failure to furnish all information required by the bidding documents or submission of a Bid not substantially responsive to the Bidding Documents in every respect would result in the rejection of the Bid.

TENDER NOTICE

INVITATION TO BID FOR THE PROVISION OF MEDICAL EQUIPMENTS UNDER THE PROJECT IMPLEMENTATION OF MASTER PLAN OF SAFARI ZOO LAHORE, PUNJAB WILDLIFE AND PARKS DEPARTMENT, LAHORE

PUNJAB WILDLIFE AND PARKS DEPARTMENT UNDER THE PROJECT IMPLEMENTATION OF MASTER PLAN OF SAFARI ZOO LAHORE TO INVITE SEALED BIDS ON SINGLE STAGE-TWO ENVELOPE BIDDING PROCEDURE FROM REPUTED FIRMS/COMPANIES. SEALED TENDER FOR PROVIDING MISC. MEDICAL EQUIPMENTS INVITED FROM AUTHORIZED DEALERS / AGENTS/ SUPPLIERS. DETAIL / DESCRIPTION OF ITEMRS IS AS UNDER.

Lot No.	Description	Qty. required	Bid Submission Date	Bid Submission Time	Bid Opening Time	Bid Document Cost(Rs)	Bid Security @ 5% of estimated cost(Rs)
Component A							
1.	Chemistry Analyzer	01	04-01-2024	10:00 AM	10:30 AM	=2000/-	5% of Estimated Cost
2.	Hematology analyzer	01		10:30 AM	11:00 AM	=2000/-	5% of Estimated Cost
3.	Microscope with camera and LED	01		11:00 AM	11:30 AM	=2000/-	5% of Estimated Cost
4.	Distillation Unit	01		11:30 AM	12:00 PM	=2000/-	5% of Estimated Cost
5.	Vortex Mixture	01		12:00 PM	12:30 PM	=2000/-	5% of Estimated Cost
6.	Hot water Bath	01		12:30 PM	01:00 PM	=2000/-	5% of Estimated Cost
7.	Misc. Lab Apparatus and Chemicals	LS		01:00 PM	01:30 PM	=2000/-	5% of Estimated Cost
8.	Hormone Analyzer	01		1:30 PM	2:00 PM	=2000/-	5% of Estimated Cost
9.	Customized V Shaped Surgical Table	01		2:00 PM	2:30 PM	=2000/-	5% of Estimated Cost
10.	Surgery Kit	01		2:30 PM	3:00 PM	=2000/-	5% of Estimated Cost
11.	Sterilization (Steamer)	01		3:00 PM	3:30 PM	=2000/-	5% of Estimated Cost
Component B							

1.	<i>Suction Apparatus</i>	01	05-01-2024	10:00 AM	10:30 AM	2000/-	5% of Estimated Cost
2.	Surgical electrical unit	01		10:30 AM	11:00 AM	=2000/-	5% of Estimated Cost
3	Gas Anesthesia Machine	01		11:00 AM	11:30 AM	=2000/-	5% of Estimated Cost
4	vital monitor	01		11:30 AM	12:00 PM	=2000/-	5% of Estimated Cost
5	Examination table	02		12:00 PM	12:30 PM	=2000/-	5% of Estimated Cost
6	Stretcher (steel)	02		12:30 PM	01:00 PM	=2000/-	5% of Estimated Cost
7	Ultrasound machine (portable)	01		01:00 PM	01:30 PM	=2000/-	5% of Estimated Cost
8	X-ray Exposure Unit	01		01:30 PM	02:00 PM	=2000/-	5% of Estimated Cost
9	Stretcher (Cloth)	04		02:00 PM	02:30 PM	=2000/-	5% of Estimated Cost
10	X-ray CR-Unit	01		02:30 PM	03:00 PM	=2000/-	5% of Estimated Cost

Last date for purchasing tender is **03-01-2024** and will be opened as per schedule given in above table.

Bidders who are not black listed from any other government department and registered with (i) Income Tax (ii) General Sales Tax Department and who are on Active Tax Payer List of Federal Board of Revenue are eligible to bid.

The Bidding document, containing detailed information, terms and conditions etc. can be obtained after publication of this notice from the office of the Project Director, Implementation of Master Plan of Safari Zoo, Punjab Wildlife and Parks Department, Lahore after payment of Rs. 2000/- for each lot through bank challan.

Bid Security as per above amount is required to be submitted along with the Technical Proposal, in the form of Call Deposit Receipt (CDR) issued by a schedule bank in favor of Project Director, Implementation of Master Plan of Safari Zoo, Punjab Wildlife and Parks Department, Lahore The bidders may bid against one lot or more than one lot separately.

Bids in a sealed single package clearly marked for Lot. No. for which bided, consisting of two sealed envelopes, clearly marked as "Technical Proposal" and "Financial Proposal" valid for a period of 90 days, from the date of opening of the bid, are required to be dropped in the tender box in the office of Project Director, Implementation of Master Plan of Safari Zoo, Punjab Wildlife and Parks Department, Lahore on the mentioned date & time.

Technical Proposals will be opened in the first instance on same day and as given above, in the office of the Project Director, Implementation of Master Plan of Safari Zoo, Punjab Wildlife and Parks Department, Lahore in the presence of the bidders or their representative who chooses to be present. The Project Director, Implementation of Master Plan of Safari Zoo, reserves the right to accept or reject the bids according to PPRA rules 2014 (amended update) after recording reasons by the committee. Bid notice is also available on PPRA website (www.ppra.punjab.gov.pk).

MUDASSER HASSAN
PROJECT DIRECTOR
IMPLEMENTATION OF MASTER PLAN OF
SAFARI ZOO LAHORE

TENDER SPECIFIC INSTRUCTION FOR BIDDERS FOR THE PROVISION OF MEDICAL EQUIPMENTS

1. Tender form can be obtained from the office of the Project Director, Implementation of Master Plan of Safari Zoo, Raiwind Road, Lahore on payment of Rs.2000/- (non-refundable) through bank challan for each lot from the date of publication of the advertisement as per directions contained in Tender Notice.
2. Tender should be dropped in the Tender box available in the office of the Project Director, Implementation of Master Plan of Safari Zoo, Raiwind Road, Lahore as per schedule given in the Tender Notice and the Technical bids will be opened on the same day at given time and date by the Tender opening committee in the presence of tenderers or their representatives, whoever to be present.
3. A bid security as per invitation of bids in favour of the Project Director, Implementation of Master Plan of Safari Zoo, Raiwind road, Lahore must accompany with tender in shape of Call Deposit Receipt (CDR) issued by any scheduled bank.
4. Bid validity of offer should not be less than 90 days from date of opening of tenders.
5. The rates should be inclusive of all applicable taxes.
6. Any erasing/cutting, overwriting will only be acceptable along with initials of bidder or clarification could be made at the time of opening in the presence of all bidders or their representative. Incomplete and ambiguous bids will not be considered.
7. Successful Tenderers will be required to execute an agreement with Project Director, Implementation of Master Plan of Safari Zoo, Raiwind Road, Lahore on stamp paper as required vide section 22(A)(b) of schedule-I of Stamp Act 1899 within (07) days after approval of the bid as per settled terms & conditions and have to furnish the performance guarantee equal to the 10% of the bid amount in the form of Bank Guarantee or Call Deposit Receipt (CDR) failing which the bid security will be forfeited.
8. Work order / Supply order will be issued to the successful tenderer after approval of the Competent Authority.
9. No escalation etc. will be admissible after the issuance of Supply order in any case.
10. The Performance Guarantee @ 10% of the total value of contract will be released after expiry of quarantine period and on the satisfactory / inspection / completion report of the concerned officer/committee.
11. The authority can cancel, all or any one of bid / Tender recording reasons in writing.
12. The number of **medical equipment's** can be reduced / increased as per availability of budget, recording reasons in writing.
13. Estimated cost of the medical equipment is 26.000 million.

Certificate:

I/We do have read all above terms and conditions and accepted as such.

Name of Tenderer: _____

CNIC No.: _____

Address: _____

Contact No.: Office: _____ Cell: _____ NTN No.: _____

Date of Tender: _____ Signature: _____

INSTRUCTIONS TO BIDDERS

INTRODUCTION

1. Scope

- a) The Purchaser wishes to receive Bid for the supply of medical Equipments mentioned in Invitation to Bid at the earlier page (Lot wise).The bid is to be completed and submitted to the Purchaser in accordance with these Instructions to Bidders.

2. Eligible Bidder

The Bidders who have duly complied with the Eligibility/Qualification and Evaluation Criteria will be eligible for consideration of financial bids.

The Bids which do not conform to the Technical Specifications or Bid conditions or the Bids from the Bidders without adequate capabilities for supply of **medical equipment's** (each lot) will be rejected.

3. Eligible Goods

- a) The medical equipment's and Services to be supplied under the Contract shall have their origin in eligible member countries.

4. Cost of Bidding

The bidder shall bear all costs associated with the preparation and submission of its Bid (s) and the Purchaser will in no case be responsible or liable for those costs.

5. Assurance

- a) The successful bidder will be required to give satisfactory assurance of its ability and intention to supply the Goods pursuant to the Contract, within the time set forth therein.
- b) The Goods shall be insured with the reputed Insurance Company at least for a period of 30 days from the date of supply.

CRITERIA FOR TECHNICAL EVALUATION OF THE TENDER / BID

(Please attach the requisite documents mentioned in the table below with the Technical Proposal)

The bid which secure 70% marks will qualify for the competition.

Mandatory Submission	Experience (40 Marks)	Managerial & Personal (HR) Capabilities (20 Marks)	Financial Position of last five years (20 Marks)	Liaison with companies/ individuals having experience of import/ export of medical equipment's (20 Marks)
<p>1. Firm registration Certificate (import/export)/ etc.</p> <p>2. Income Tax Returns of Last Two Years</p> <p>3. Affidavit on Stamp paper that the bidder is not blacklisted by the Government or any of its department / institutions at any forum</p> <p>4. Active Taxpayer Certificates / Evidence under the Income Tax & Sale Tax Authority in Current date</p> <p>5. Certificate of Punjab Revenue Authority</p>	<p>40 marks for having three or more similar contracts.</p> <p>30 marks for having two similar contracts.</p> <p>20 marks for having 1 similar contract.</p> <p>No marks for having no similar project.</p> <p>(Attach proof / bill of landing/ online inquiry etc.)</p>	<p>Management Staff.</p> <p>1. 01 Professional Qualified Manager having 16-year education with 05-year experience (05 Marks)</p> <p>Other Staff (05 Marks)</p> <p>2. Company Profile (10 marks)</p>	<p>20 marks for annual cash flow of last 05 years of the firm is equal to the estimated cost.</p> <p>If low following formula will be used to calculate the numbers:</p> <p>(A/X) x20 A= turnover X= estimated cost</p> <p>(Attach bank statement)</p>	<p>20 Marks for liaison with two companies</p> <p>10 marks for liaison with one company</p> <p>No marks for liaison with no company</p> <p>(Attach proof / collaboration certificate etc.)</p>
(a)	(b)	(c)	(d)	(e)

BIDDING DOCUMENTS (Each Lot)

6. Contents of Bidding Documents

- 6.1 The goods required, bidding procedures and Contract terms are prescribed in the bidding documents. In addition to the Invitation for Bid, the bidding documents include:
- a) Instructions to Bidders
 - b) Technical Specifications
 - c) Bid Form
 - d) Schedules:
 - Schedule - A Special Stipulation
 - Schedule - B Price Schedule
 - e) Bid Security Form
 - f) Performance Security Form
 - g) Conditions of Contract
- 6.2 The bidder is expected to examine all instructions, forms, terms and specifications in the Bidding Documents. Failure to furnish all information required by the Bidding documents or submission of a bid not substantially responsive to the bidding documents in every respect will result in the rejection of the bid.

7. Clarification of Bidding Documents

- 7.1 The prospective bidder requiring any further information or clarification of the bidding documents may notify the Purchaser in writing or by visiting at the following address:
Project Director,
Implementation of Master Plan of Safari Zoo, Raiwind Road, Lahore or his nominated person will respond in writing to any request for information or clarification of the bidding documents which it receives no later than three (03) days prior to the deadline for the submission of bid.

8. Amendment of Bidding Documents

- 8.1 At any time prior to the deadline for submission of bid, the Purchaser may, for any reason, whether at its own initiative or in response to a clarification requested by the prospective bidder, modify the bidding documents by amendment.
- 8.2 The amendment shall be part of the bidding documents, pursuant to Clause 7.1, will be notified in writing or by e-mail or by fax to the prospective bidder who has received the bidding documents, and will be binding on him. Bidder is required to acknowledge receipt of any such amendment to the bidding documents.
- 8.3 In order to afford the prospective bidder reasonable time in which to take the amendment into account in preparing its bid, the Purchaser may, at its discretion, extend the deadline for the submission of bid but in the manner similar to the original advertisement.

PREPARATION OF BID

9. Language of Bid

The bid prepared by the bidder and all correspondence and documents relating to the bid exchanged by the bidder and the Purchaser shall be written in the English / Urdu language. Any printed literature furnished by the bidder may be written in another language, provided that this literature is accompanied by an English / Urdu translation, in which case, for purposes of interpretation of the bid, the English / Urdu translation shall govern.

10. Documents Comprising the Bid

The evaluation of Bid submitted shall be inclusive of, but not be limited to, the following factors:

a) Bid Form

The Bidder shall complete the Bid Form in accordance with Clause 11.

b) Price Schedule

The Bidder shall complete the appropriate Price Schedule furnished in the Bidding Documents in accordance with Clauses 12 & 13.

c) Bid Security

The bidder shall furnish Bid Security in accordance with Clause 14.

11. Bid Form

The bidder shall complete the Bid Form duly signed by the authorized personnel along with the stamp of the company and all the Schedules furnished in the bidding documents.

12. Bid Prices

12.1 The bidder shall complete Schedule B, Price Schedule for all Items / Equipment on which he wants to quote rate as per the instructions contained in this document.

12.2 Prices quoted in the Price Schedule for the equipments should be entered in the following manner:

(i) The price of the Goods will be quoted (for) Lahore inclusive of appropriate transportation at the address provided in Schedule (A).

(ii) Bidders shall quote rate and Technical Specifications for all equipment's as per unit mentioned in Price Bid Schedule (B).

(iii) The blank or partially / conditionally filled Price Bid Schedule B of any item is considered noncompetitive.

(iv) The price is to be submitted in Pak Rupees only and should include all taxes.

12.3 Prices quoted by the bidder shall remain fixed and valid until completion of the Contract performance and will not be subject to variation on account of escalation.

13. Bid Currencies

Prices shall be quoted in Pak Rupees.

14. Bid Security

14.1 Pursuant to Clause 10.1(c), the bidder shall furnish, as part of its bid, a bid security in the amount equal to mentioned in invitation of bids in the form of CDR.

14.2 The bid security shall be denominated in Pak Rupees and shall be in shape of call Deposit Receipt (CDR) in favor of the Project Director, Implementation of Master Plan of Safari Zoo, Raiwind Road, Lahore.

14.3 The bid not secured in accordance with Clauses 14.1 and 14.2 above will be rejected by the Purchaser as non-responsive.

14.4 Bid security of un-successful Bidder's will be discharged or returned, or both, as promptly as possible upon award of Contract.

- 14.5 The successful bidder's bid security will be returned, upon executing the contract, pursuant to Clause 29, and furnishing the performance guarantee, pursuant to Clause 30.
- 14.6 The bid security may be forfeited:
- (a) If the bidder withdraws its bid during the period of bid validity specified by the bidder on the Bid Form; or
 - (b) If the bidder does not accept the correction of the Total Bid Price pursuant to Clause 22; or
 - (c) If the bidder fails;
 - (i) To sign the contract in accordance with Clause 29, or
 - (ii) To furnish the performance guarantee in accordance with Clause 30.
- 15. Period of Validity of Bid**
- 15.1 The bid shall remain valid for ninety (90) days from the date of bid closing prescribed by the Purchaser.
- 15.2 Notwithstanding Clause 15.1 above, the Purchaser may solicit the bidder's consent to an extension of the period of bid validity. The request and the responses thereto shall be made in writing or by cable or fax. If the bidder agrees to the extension request, the validity of the bid security provided under Clause 14 shall also be suitably extended. The bidder may refuse the request without forfeiting its bid security. The bidder will not be required or permitted to modify its bid.
- 16. Format and Signing of Bid**
- 16.1 The bidder shall submit duly filled original Bidding Document issued to him.
- 16.2 Prescribed Bid Form and Schedules shall be used and not to be retyped. The original bid shall be typed or written in indelible ink and shall be signed by the bidder or a person duly authorized to sign on behalf of the bidder. Such authorization shall be indicated by written power of attorney accompanying the bid. All pages of the bid where entries and amendments have been made shall be initialed by the person signing the bid.
- 16.3 The bid shall contain no interlineations, erasures or overwriting except as necessary to correct errors made by the bidder, in which case such corrections shall be initialed by the person or persons signing the bid.

SUBMISSION OF BID

17. Sealing and Marking of Bid

17.1 The bidder shall seal the original Bid in an envelope.

17.2 The envelopes shall:

(a) Be addressed to the following address:

The **Project Director**,
Implementation of Master Plan of Safari Zoo Lahore,
Raiwind Road, Lahore.

(b) Bear the following identification:

Bid for “**Provision of Medical Equipment’s (Lot No.)**”

DO NOT OPEN BEFORE given time on 04-01-2024 and 05-01-2024 for respective component.

Note: Opening time will be as per time mentioned in invitation of bid(s)

17.3 In addition the envelope shall indicate the name and address of the bidder to enable the bid to be returned unopened in case it is declared “Late”.

17.4 Fax bid will not be entertained.

18. Deadline for Submission of Bid

The original bid must be submitted by the bidder at the office of the Project Director, Implementation of master plan of Safari Zoo Lahore, Raiwind Road, Lahore by given time and date.

19. Late Bid

The bid received by the Purchaser after the deadline for submission of bid prescribed by the Purchaser, pursuant to Clause 18, will be rejected and returned unopened to the bidder.

OPENING AND EVALUATION OF BID

20. Opening of Bid

- 20.1 The bid shall be opened by the Purchaser in the presence of the bidder's representatives who choose to attend at the time and date specified in Clause 18, at the office of the Purchaser, given in Clause 17.2 (a). The bidder's representatives who are present shall sign an attendance sheet evidencing their attendance.
- 20.2 The bidder's name, bid price, modifications and the presence or absence of the requisite bid security, and such other details as the Purchaser, at its discretion, may consider appropriate will be announced and recorded at the opening.

21. Clarification of Bid

To assist in the examination, evaluation and comparison of bid, the Purchaser may, at its discretion, ask the bidder for a clarification of its bid. All responses to requests for clarification shall be in writing, and no change in the price or substance of the bid shall be sought, offered or permitted except as required to confirm the corrections of arithmetical errors discovered in the evaluation of bid.

22. Determination of Responsiveness of Bid

- 22.1 Prior to the detailed evaluation of the bid, pursuant to Clause 26, the Purchaser will examine and determine the substantial responsiveness of the bid to the requirements of the bidding documents. A substantially responsive bid is one which:
- (a) Meets the eligibility criteria specified in Clauses 2 and 3;
 - (b) Has been properly signed on the Bid Form;
 - (c) Is accompanied by the required Securities and these Securities are valid and in good order;
 - (d) The technical specifications for **of Medical Equipments (each lot)** quoted in the price Bid Schedule B should meet the major technical criteria as specified for **Medical Equipments (each lot)** in technical specifications of this document;
 - (e) Meets the delivery period set out in Schedule A Special Stipulations to Bid;
 - (f) Meets the rate and limit of liquidated damages as specified in Schedule A Special Stipulations to Bid;
 - (g) Offers fixed price quotations i.e. the bid do not offer an escapable price quotation;
 - (h) Is otherwise complete and generally in order;
 - (j) Conforms to all the terms, conditions and Specifications of the bidding documents, without material deviation or reservation. A material deviation or reservation is one:
 - (i) Which affects in any substantial way the scope, quality or performance of the Goods or
 - (ii) Which limits in any substantial way, inconsistent with the bidding documents, the Purchaser's rights or the bidder's obligations under the Contract.
- 22.2 The bidder's responsiveness shall be based on the contents of the bid itself without recourse to extrinsic evidence.

- 22.3 The bid determined as not substantially responsive will be rejected by the Purchaser and may not subsequently be made responsive by the bidder by correction or withdrawal of the nonconforming deviation or reservation.
- 22.4 The Purchaser may waive any minor informality or non-conformity or irregularity in the bid.
- 22.5 Correction of Arithmetical Errors: Bid determined to be substantially responsive will be checked by the Purchaser for any arithmetic errors. Errors will be rectified as follows:
- (a) for the total individual bid price of each item entered in paragraph of the Bid Form, if there is a discrepancy between the amounts in Figures and in words, the amount which tallies with the total individual Bid Price of each item, shown in the Price Schedule for each item, will govern unless the Bid purchase (committee) contains a specific statement confirming the total individual Bid Price of each item.
 - (b) where there is a discrepancy between the unit rate and the total price resulting from multiplying the unit rate by the quantity, the unit rate as quoted will govern and the total price shall be corrected, unless in the opinion of the Purchaser, there is an obviously gross misplacement of the decimal point in the unit rate, in which case the total price for each item as quoted will govern and the unit rate will be corrected, and
 - (c) Where there is a discrepancy in the total price quoted in the Price Schedule of each item / equipment vis-à-vis addition of each item, the total of the itemized prices will govern.

The amount stated in the Bid Form will be adjusted by the Purchaser in accordance with the above procedure for the correction of errors, and shall be considered as binding upon the Bidder. If the Bidder does not accept the correction of the errors for any item in the Bid, his Bid will be rejected for the specific item and the Bid Security for that item will be forfeited in accordance with the Clause 14.6.

Corrected Total Individual Bid Price for each item the price as determined after the application of arithmetic corrections shall be termed as Corrected Total individual Bid Price for each item.

23. Evaluation and Comparison of Bids

The Purchaser will evaluate and compare the bids previously determined to be substantially responsive, pursuant to Clause 22, as stated hereinafter.

23.1. Basis of Evaluation and Comparison of Bid

The Bids of only those Bidders who are substantially responsive to the requirements of the Bidding document will be considered for evaluation. The evaluation and comparison of the Bid will be done total in value. The offer/ contract will be awarded to those Bidders who's corrected and evaluated total Bid Price is the lowest.

23.2. Evaluated Bid Prices

The Purchaser evaluation of a bid will take into account in addition to the Bid Price, the following factors (Adjustments) in the manner and to the extent stated hereinafter. Adjustment will be based on corrected Bid Prices. The price so determined after making such adjustments will be termed as Evaluated Bid Price. Correction of arithmetical errors as stated in Clause 22.5. The cost of making good any deficiency resulting from any acceptable, quantifiable variations and deviations from the Schedules, Conditions

of Contract and Specifications, shall be added to the corrected Bid Price for comparison purposes only. For bid offering delivery period of the Service earlier than the period specified in the Schedule A Special Stipulations to Bid, no credit will be given.

Terms of Payment: The bidder shall state their bid price for the payment terms outlined in the Conditions of Contract. The bid will be evaluated on the basis of this base price. The bidder may state alternate payment terms and indicate the reduction in bid price that wish to offer for such alternative payment terms. The Purchaser may consider the alternative payment terms offered by the bidder.

24. Contacting the Purchaser

Any effort by a bidder to influence the Purchaser in the Purchaser's decisions in respect of bid evaluation, or Contract award will result in the rejection of the bidder's bid.

25. Purchaser's Right to Accept the Bid or Reject the Bid

The Purchaser reserves the right to accept or reject any bid and to cancel the bidding process at any time prior to award of Contract, without thereby incurring any liability to the bidder for the Purchaser's action as per PPRA rules 2014 (amended up to date).

AWARD OF CONTRACT

26. Post-qualification and Award Criteria

- 26.1 The Purchaser will determine to its satisfaction whether the bidder has offered Service at reasonable prices consistent with the current prevailing market prices and is qualified to satisfactorily perform the Contract.
- 26.2 An affirmative determination will be prerequisite for award of the Contract to the bidder. A negative determination will result in rejection of the bidder's Bid.
- 26.3 Subject to Clause 26 above, the Purchaser will award the Contract to the bidder if its bid has been determined to be substantially responsive to the bidding documents and consistent with the current prevailing market prices as determined by the Purchaser, provided further that the bidder is determined to be qualified to satisfactorily perform the Contract.

27. Purchaser's Right to Vary Quantities at Time of Award

The Purchaser reserves the right at the time of award of Contract to increase or decrease the quantity of goods specified in the Specifications upto a reasonable level without any change in unit prices or other terms and conditions.

28. Notification of Contract Award

Prior to the expiration of the period of bid validity, the Purchaser will notify the bidder in writing by registered letter that its bid has been accepted. This letter is termed as Letter of Acceptance.

29. Signing of Contract

- 29.1 After the acceptance of bid, the Purchaser will send to the successful bidder the Contract Form provided in the bidding documents, incorporating all agreements between the parties.
- 29.2 Within **five (5) days** of the receipt of such Contract Form, the bidder shall sign and date the Contract and return it to the Purchaser.

30. Performance Guarantee

Within **seven (7) days** of the receipt of the Letter of Acceptance from the Purchaser, the bidder shall furnish the performance guarantee, in accordance with the Conditions of Contract, in the performance guarantee Form provided in the bidding documents or another form acceptable to the Purchaser.

ADDITIONAL INSTRUCTIONS

31. Instructions to Assist the Bidder

Bid shall be prepared and submitted in accordance with the instructions set forth herein. These instructions to Bidders are provided to assist in preparing their Bid and shall not constitute part of the Contract Documents.

32. Income Tax & General Sales Tax

The bidder may make inquires on income tax to the concerned authorities of Income Tax and General Sales Tax Department, Government of Pakistan.

BID FORM (LOT NO.1)

(To be submitted with Financial Proposal)

To:

The Project Director,
Implementation of Master Plan of Safari Zoo Lahore,
Raiwind Road, Lahore.

Gentlemen:

Having examined the bidding documents including Addendum No _____, the receipt of which is hereby duly acknowledged, for the above Contract, we, the undersigned, offer to supply, deliver test and impart training in conformity with the said bidding documents for the Total Bid Price for items / equipment detailed below:

Lot No.	Description	Qty. required	Lot No.	Description	Qty
<i>Component A</i>					
1.	Chemistry Analyzer	01			
2.	Hematology analyzer	01			
3.	Microscope with camera and LED	01			
4.	Distillation Unit	01			
5.	Vortex Mixture	01			
6.	Hot water Bath	01			
7.	Misc. Lab Apparatus and Chemicals	LS			
8.	Hormone Analyzer	01			
9.	Customized V Shaped Surgical Table	01			
10.	Surgery Kit	01			
11.	Sterilization (Steamer)	01			
<i>Component B</i>					

1.	<i>Suction Apparatus</i>	01			
2.	Surgical electrical unit	01			
3.	Gas Anesthesia Machine	01			
4	vital monitor	01			
5	Examination table	02			
6	Stretcher (steel)	02			
7	Ultrasound machine (portable)	01			
8	X-ray Exposure Unit	01			
9	Stretcher (Cloth)	04			
10	X-ray CR-Unit	01			

(Please fill)

We undertake, if our above stated total Bid for medical Equipments (Lot No.) is accepted, to complete the Works in accordance with the Contract Execution Schedule provided in the Schedule A Special Stipulations to Bid.

We will provide the performance guarantee in the sum equivalent to 10 % of the Contract Price, for the due performance of the Contract.

We agree to abide by this Bid for the period of Ninety (90) days from the date fixed for bid opening pursuant to Clause 22 of the Instructions to Bidders, and it shall remain binding upon us and may be accepted at any time before the expiration of that period.

Until a formal Contract is prepared and executed, this Bid, together with your written acceptance thereof in your Notification of Contract Award, shall constitute a binding Contract between us.

We understand that you are not bound to accept the lowest-price or any Bid that you may receive.

Dated: _____

Bidder Signature Along With Seal

WITNESS
Signature

Name

Title

Address

Signature

Name

Title

Address

SCHEDULE –A: SPECIAL STIPULATIONS

(To be Submitted with Technical Proposal)

For ease of reference, certain information and Special Stipulations applicable to the Contract are set forth herein.

1	Amount of Performance Guarantee	Clause 30 of instructions to Contractor.	The selected firm will have to deposit 10% Performance Guarantee in the form of Call Deposit Receipt (CDR) at the time of Letter of award. This amount will be returned to the firm up to 60 days after Contract completion.
2	Delivery of Good / Services		
	Address at which Goods / Equipment's to be delivered	---	Project Director, Implementation of Master Plan of Safari Zoo Lahore, Raiwind Road, Lahore or the location specified by this office.
3	Contract Execution Schedule.	---	After issuance of Purchase order
3	Start operation	---	Just after issuance of taking over certificate
4	Work completion		Within 90 days
5	Payments		
	Time within which payment shall be made after the delivery of certificate of payment to purchaser and inspection of work done by the committee	---	Payment will be made after successful completion of awarded work within stipulated period and satisfactory report from inspection committee.
6	a) Liquidate damages for delayed delivery thereof. b) Limit of Liquidated Damages		01% of the Contract Price of the delayed per day or part thereof, if the extension in work order will not be granted by competent authority. Not to exceed in the aggregate ten (10%) of Contract Price stated in the Notice of Contract Award/Letter of Acceptance by the Purchaser.

Signature of the Bidder: _____

SCHEDULE - B: PRICE SCHEDULE

(To be Submitted with Financial Proposal)
DETAIL OF MEDICAL EQUIPMENTS (LOT WISE)

Lot No.	Description	Qty.	Rate each	Total Amount
<i>Component A</i>				
1.	Chemistry Analyzer	01		
2.	Hematology analyzer	01		
3.	Microscope with camera and LED	01		
4.	Distillation Unit	01		
5.	Vortex Mixture	01		
6.	Hot water Bath	01		
7.	Misc. Lab Apparatus and Chemicals	LS		
8.	Hormone Analyzer	01		
9.	Customized V Shaped Surgical Table	01		
10.	Surgery Kit	01		
11.	Sterilization (Steamer)	01		
<i>Component B</i>				
1.	<i>Suction Apparatus</i>	01		
2.	Surgical electrical unit	01		
3.	Gas Anesthesia Machine	01		
4	Vital monitor	01		

5	Examination table	02		
6	Stretcher (Steel)	02		
7	Ultrasound Machine (Portable)	01		
8	X-ray Exposure Unit	01		
9	Stretcher (Cloth)	04		
10	X-ray CR Unit	01		

**Rates must be inclusive of all applicable taxes.
Required Specifications are attached at Page no. 25 and 26**

Total Price in Words (Inclusive of All Taxes):

Signature of Bidder _____

**TERMS OF REFERENCE
FOR
SUPPLY OF MEDICAL EQUIPMENTS**

Scope of work:

Supply of medical equipment's (Each Lot) according to Specification.

Functional Requirements and Device Specifications

1. The firm should visit the office of the Project Director, Implementation of Master Plan of Safari Zoo, Raiwind Road, Lahore in order to gauge the user requirements.
2. The firm will be responsible to do the procurement / Supply of items / complete the job.
3. The description / specifications of items to be purchased have been mentioned as under: -

SPECIFICATIONS / DESCRIPTION

COMPONENT A

1. Chemistry Analyzer

Specific for veterinary use, Sample type (whole blood or plasma or serum, Sample volume (100 nl or less), Minimum parameter 28 or more (Especially including Alt, AST, CX, BUN, Ca, P, glucose, K, Na, Na/p, CO₂, Mg, Bilirubin). Could support multiple species especially Elephants, With properly installed software PC (if required) and Printer. With Rotor Technology. 10 rotors of RFTs, 10 Rotors of LFTs and 10 rotors of basic electrolytes will be supplied with machine.

Power 100 to 240 V/AC/Single phase.

Made in: USA/ UK/ Germany/ Japan/ Canada/ China (Good Quality).

2. Hematology Analyzer

Specific for veterinary use, Autohemology analyzer, could perform CBC, 4 part differential especially eosinophil number and percentage. Could show eosinophils graphs. Open vial sampling system, minimum blood sample: 10ul or

less, More than 50000 Result storage capacity, Could measure 23 parameter or more, Could support more than 19 animal species and have facility to add more animals profile, Touch screen display

Must have USB ports (for External Printer,Wi-Fi, keyboard) and LAN-port

Built in printer and one external printer (also PC if required). Properly installed software .With all accessories and one set of all reagents.

Power 100 to 240 V/AC/Single phase

Made in: USA/ UK/ Germany/ Japan/ Canada/ China (Good Quality).

3. Microscope With Camera & LED

Infinity Optical System, Transmitted microcopy with LED illumination, trinocular tube FN20 (Minimum), pair of wide field eye piece 10X with helicoid adjustment, 360 Rotatable eye piece. Quadruple revolving nose piece, double plat stage, abbe condenser, and plan achromatic objectives 4X, 10X, 40X, 100X oil

Coarse & fine focus knobs, condenser NA 1.25 or better, moveable stage height, coarse adjustment limit stopper.

Complete accessories, Box, dust Cover, adapters, Power cords etc

Power: AC/ Single phase/ 210 - 240 Volts

Country of Origin/ Manufacture: USA/ UK/ Germany/ Japan/ Canada.

Microscope Digital Camera

HDMI C-Mount Camera

Minimum 6mp still resolution on 3264 x 1836 or better.

Minimum 2mp live video 1920 x 1080 60fps Frames or better via HDMI

Minimum 2MP live video 1920 x 1080 30fps or better via USB

Output: HDMI, USB, SD CARD, or More.

Minimum Controls: white Balance, Gamma, gain, sharpness, saturation, cross line image compare, Image freeze.

With all installed Software.

Compatible with Windows XP and later.

Power: AC/ Single phase/ 210 - 240 Volts

Country of Origin / Manufacture (Camera): USA/ UK/ Germany/ Japan/ Canada/ Europe.

32` LED (Good quality) compatible with Camera & Microscope

4. Distillation Unit

Borosilicate Glass Body

Minimum Capacity 4L per Hour

Metal Element: Minimum 3 KW

PH 5.5 ---- 6.5

Country of Origin: USA/ UK/ Malaysia/ Japan/ Canada.

5. Vortex Mixer

Speed: 0 – 2500 rpm

Run type: Touch Operational/ Continuous with all Adapters for different sized Lab tubes.

With complete accessories and power cords.

Power: AC/ Single phase/ 210 - 240 Volts

Made in: USA, UK, Canada, Germany, China (Good Quality), Europe

6. Hot Water Bath

Stainless steel / Corrosion resistant body with 4 holes.

Reservoir capacity: 10L Minimum

MAX temperature: 99.9° C or more

T. Stability $\pm 0.5^{\circ}$ C or less

With Drain valve, Lid on holes, LED Digital temperature display which should show actual temperature & set temperature.

Low water level and over temperature control.

With all standard accessories and power cords.

Power: AC/ Single phase/ 210 - 240 volt.

Made in: USA/ UK/ Germany/ China/ Europe.

7. Misc. Lab Apparatus & Chemicals

Sr.No	Name of Item	Quantity
1	M _c master slide	2
2	CBC slide	2
3	Glass Pipette ---10ml	4
4	Beaker 1 litre	4
5	Lab thermometer	5
6	Beaker tongue	2
7	Glass slides – 5box	5 box
8	Corer slides- 5box	5 box
9	Lab sample transportation insulation Box	2
10	Head caps	1000
11	Disposable gowns	30
12	Decent sheets	30
13	Safety goggles	5
14	Micropipette 10 to 100 ul	1
15	tips- yellow color	500
16	Micropipette 100 ul to 1000ul	1
17	Blood sampling vacutainer (EDTA added)	200
18	Blood sampling vacutainer (with coagulation activator)	200
19	Plane blood sampling vacutainers	200
20	Polythene zipper different	1 kg
21	Giemsa stain 450 ml	4

22	95 alcohol	1 litre
23	KOH	1 litre
24	Centrifuge Tube Plastic 10ml	30
25	Centrifuge Tube Plastic 5ml	30
26	Micro Centrifuge tube	100
27	Field stain A 450 ml	4
28	Field stain B 450 ml	4
29	Cedar wood oil 500ml	2
30	Sodium citrate solution –450 ml (for fecal parasite test)	4
31	HB reagent N/10 450ml	1
32	WBC reagent 450ml	1
33	Platlets reagent 450ml	1
34	WBC chamber	2
35	Blue tips for micropipette	500
36	Flask- 500ml	4
37	Vicryl 1-0	20
38	Vicryl 2-0	20
39	Vicryl 1	20
40	Vicryl 2	20
41	Vicryl 3	20
42	Silk Suture 1-0	20
43	Silk suture 2-0	20
44	Silk surure 1	20

45	Silk suture 2	20
46	Silk suture 3	20
47	Methylated spirit	2 litre
48	Benzene 100ml	1

8. Hormone Analyzer (Vet Immunofluorescence Quantities Analyzer)

Specific for veterinary use, Sample Type: whole blood, serum or plasma, Built in thermal printer, fully installed software, LCD-touch screen, Reaction time 15 min or less could detect hormones, special markers, inflammation factors. With at least 20 T4 and 20 Progesterone cards/kits, support Wi-Fi data transmission, with built in temperature control. With r sets of printer paper and all accessories. Power 100 to 240 V/AC/Single phase

Made in: USA/ UK/ Germany/ Japan/ Canada/ China (Good Quality).

9. Custom V. Shaped Surgical Table

Two pieces stainless steel platform longitudinally attached which could be folded. Minimum width 30 inches. Minimum length 66 inches. Stainless steel sheet should be 18 gauge minimum. Drain under slit b/w two pieces of platform which should lead to wastebin. Hydraulic control to adjust height. Platform should be **heat able** (with good current & Temperature safety), with bedding (minimum 2 inch Foam bedding covered with blue/black water proof leather/ Rexene cover). Four hooks under platform to tie animal. Made in Pakistan.

10.SURGERY KIT

Sr.No	Name of Item	Quantity
1	Scalpel handles 03 No.	3
2	Scalpel handles 04 No.	3
3	Mayo scissor curved Large (approx. 8 inches)	3

4	Mayo scissor Straight Large (Approx. 8 inches)	3
5	Mayo scissor (straight) small (approx. 5 inches)	3
6	Mosquito forceps	5
7	Thumb forceps with teeth	5
8	Thumb forceps without teeth	5
9	Artery forceps	10
10	Towel clamps	30
11	Cobb elevator	01
12	Carmalt forceps	05
13	Bone curette	02
14	Cottle angular scissor	03
15	Surgical tray large	05
16	Surgical tray small	05
17	Kidney tray large	05
18	Surgical blade 12 No	50
19	Surgical blade 23 No	50
20	Surgical blade 24 no	50
21	Kelly clamp	06

22	Tonsil hemostat	06
23	K wire cutter	02
24	Intestinal forceps	04
25	Hand held retractor	04
26	Periosteal Elevators (round edge, straight edge)	02
27	Orthopedic drill and different size bits	01
28	Wire twister	01
29	Hoof knife	2
30	Hoof disc	2
31	Baby grinder for hoof disc	1
32	Orthopedic Chisel, Mallet	2
33	Steinmann pins (Different Size)	20
34	Bone cutter	02
35	mayo Hagar needle holder	06
36	Metzenbam scissor	12
37	Instrument tray for surgery	02
38	Pig holder instrument	01
39	Sponge forceps	04

40	Vaginal speculum large	01
41	Addison's forceps.	06
42	Allis tissue forceps.	06
43	Steel knife.	02
44	Trocar + cannula.	02
45	Rongeurs.	02
46	Jacobs chuck and key.	01
47	Internalsplints/bone plates with screws (7 bits)	4
48	Internalsplints/bone plates with screws (9 bits)	4
49	Instrument Trolley for operation theatre	1
50	Dental scaler along with adapter and accessories	1
51	Tips for Dental Scaler	5

11. Sterilization Steamer

Automatic bench top /desktop/table top pressure autoclave steam sterilizer.

Made of stainless steel. With ultrahigh temperature alarm and temperature control, LCD Display, CE and ISO Certified. Temperature range 105 to 134 celciusor wider. Volume 24 L minimum.

With all accessories (Internal trays, clippers, Power cords etc)

Power 100 to 240 V/AC/Single phase

1. Country Origin: USA/ UK/ Germany/ Japan/ Canada/ China (Good Quality)

COMPONENT B

1. Surgical Suction Apparatus

With trolley, suction tubes, two 5liter lid covered bottles jars which should be graduated & autoclavable, overflow protection, Change over valve system, all other accessories and power cables.

Power: AC power / Single phase/ 210 - 240 volt

Made in: USA/ UK/ Germany/ Europe/ China (Good Quality).

2. Electrosurgical Unit / (Diathermy Unit)

With double foot paddle cutting & coagulation, handle controlled pencil, Bipolar forceps with cable, Patient plate & cable, Bipolar foot paddle, Return plate cable, Cutting forceps, complete accessories and power cables.

Power: AC power / Single phase/ 210 - 240 volt

Made in: Pakistan/USA/ UK/ Germany/ China/ Europe/ Canada.

3. Gas Anesthesia Machine

Specific for veterinary use which could support anesthesia of African lion and zoo animal of its size.

Ventilator

Electronic turbine operated, Modes: Volume Support (VS) Volume Control Ventilation (VCV), Pressure Control Ventilation (PCV) , Synchronized Intermittent Mandatory Ventilation SIMV) or more, Touch screen with brightness control.

Having option of CO2 Monitoring.

Integrated breathing system.

Vaporizer: High pressure vaporizer with Pour Fill, Key Fill/Quick-Fill method. Isoflurine supported.

Air and oxygen type pneumatic system with proper pipelines.

Waste gas scavenging system (passive as well as active) with waste gas canister and weight monitor (weight range 0g to 2 kg or more).

With **oxygen generator/cylinder** with its fitting/tray.

With trolley, anesthesia gas filter canister, and all standard accessories.

Chargeable along with lithium battery minimum 5000 mah, Minimum run time 120 minutes or better.

Isoflurane and all other required chemicals, pipes, endotracheal tubes for deer and lion will be provided with machine so that it may be ready to use. With fully installed **software**.

Multi-functional Communication Port for communication between the anesthesia machine and external devices. SB / Network port for software upgradation or transfer of data

Made in: USA/UK/Germany/Japan/Korea/Canada/China (Good Quality).

Specific for vet use along with ventilator trolley, oxygen cylinder and complete accessory, Display touch screen with brightness control

Chargeable along with battery ,Minimum battery time 90min, Vaporizer, Gas type O₂ or Air

4. Vital Monitor

Specific for veterinary use

High quality body and accessories ,Touch screen (Minimum 12 inches) , Portable, display screen, Weight less than 5 kg, Chargeable, Basic parameters (ECG, SPO₂, NIBP, Pulse Rate, Respiration Rate, Body temperature), Data storage function, 5 ECG leads set , ECG Accuracy +-0.02Mv or better, Heart rate range 15 to 340bpm or wider range with resolution of 1 bpm, Respiration range:0 to 150 or wider range, SpO₂ Range 0 to 100 percent with resolution of 1 percent, Pulse rate 20 to 250 or wider range. Supported for barcode scanner and wifi. Having network connector interface, USB Connector interface, VGA output connector interface and output connector interface for ECG and nurse call. Chargeable with more than 2300 mAh capacity and 3.5 hr run time. Could measure systolic, diastolic and mean NIBP.

Power 100 to 240 V/AC/Single phase.

Made in: USA/ UK/ Germany/ Japan/ Canada/ China (Good Quality).

5. Examination Table

Width: 24 inches, Length: 60 inches, Minimum height: 36 inches

With drain hole, pipe & stainless steel lid covered wastebin.

18 gauge non corrosive stainless steel top with 2 inches raised edges.

Frame of 18 gauge galvanized / stainless steel pipe

Made in Pakistan.

6. Stretcher (Steel)

Length: 66 inches, Width: 27 inches

Frame of 16 gauge galvanized / stainless steel pipe with handles.

18 gauge non corrosive stainless steel top, deep from center.

Three belts to hold animal. With bedding (minimum 2 inch Foam bedding covered with blue/black water proof leather/ Rexene cover).

With Trolley: To carry this stretcher. Frame of trolley should be made of 16 gauge galvanized iron pipe. Tyres of trolley: solid, with locks, durable, minimum diameter 5 inches.

Made in Pakistan

7. Ultrasound Machine

Specific for veterinary use, with Multi-frequency Convex Probe (minimum 3.5 to 5 MHz) , Multi-frequency Micro Convex Probe (Minimum 5.5 to 9MHz), Multi-frequency Rectal Probe (minimum 4.5 to 8 MHz)

System Transducer Frequency: 2.0-10.0 MHz or wider range, Minimum Display 12.1" TFT, Cine Loop Memory 256 frames or better, Zoom: 8 steps or better, Display Modes: B, 2B, 4B, B+M, M, and PW or more, Frame rate: 250 f/s or better, Scanning Depth: 320mm or better, Minimum applications Large Animal: Reproduction, Tendon. Small animal: Abdomen, Small part, basic cardiology. internal large-capacity data storage, Built-in image archive: 504MB built-in image storage or more

With USB port: VGA output, Minimum Two probe connectors , Built-in Battery with minimum 90 minutes back-up operation , with all accessories(power cable, 2 gel bottles) and properly installed software for canine, bovine, feline, equine and ovine.

Power 100 to 240 V/AC/Single phase

Country Origin: USA/ UK/ Germany/ Japan/ Canada/ China (Good Quality).

8. X-RAY Exposure Unit

Portable X-Ray unit, less than 5 kg in weight, minimum 5 MA power, built in generator with less than 500 Watt power, Anode X-Ray tube with upto 10 KJ heat capacity. Equipped with collimator. Could take clear chest X-ray of deer, Chargeable, Battery backup: could shoot Up to 100 or more images on Single charge compatible With CR and DR, Could Provide high contrast images without

using GRID, Should be supplied with complete working accessories, Shoulder strap, Lead apron and supporting stand. Power 100 to 240 V/AC/Single phase

Made in: USA/ UK/ Germany/ Japan/ Canada/ Korea.

9. Stretcher (Cloth)

Made of good quality canvass cloth with longitudinal 16 gauge galvanized iron pipe handles and three belts.

Two of them

Length: 66 inches, Width: 42 inches

Two of them

Length: 40 inches, Width: 24 inches,

Made in Pakistan

10.CR Unit (X-RAY Image Reader)

Computed Radiography X ray Reader, Supported with Veterinary software Processing capacity 70 IPs/hour, Reading specification 10 pixel/mm or better, time to start on display : less than 40 sec, in table top design . With three 14*14 IP cassette and three 14*17 inches IP Cassette. Compatible with multiple x ray exposure units.

With Installed software, computer, 3KW ups (sign Wave online) , Power cords and standard accessories.

Power 100 to 240 V/AC/Single phase

Country Origin: USA/ UK/ Germany/ Japan/ Canada.

BID SECURITY FORM (LOT NO.1)

(To be Submitted with Financial Proposal)

The Total Bid Security amounting to Rs._____ in shape of “Call Deposit Receipt (CDR)” of the Bank (Name)_____ is attached in accordance with Clause 14 of the Instruction to Bidder for the supply of following Items.

Core Requirements

Sr. No.	Description	Amount (Rs.)
1	Provision of Medical Equipment's	
Total Amount (Inclusive of all taxes) LOT NO.		

Signature of Bidder _____

PERFORMANCE GUARANTEE FORM (LOT NO. 1)

To:

The Project Director,
Implementation of Master Plan of Safari Zoo Lahore,
Raiwnd Road, Lahore

WHEREAS (Name of the Contractor)

Hereinafter called "the Contractor" has undertaken, in pursuance of "Invitation to Bids / Short Tender", Provision of medical equipment's (Lot No.) (Hereinafter called "the Contract").

AND WHEREAS it has been stipulated by you in the Contract that the Contractor shall furnish you with a bank guarantee by a recognized bank for the sum specified therein as security for compliance with the Contractor's performance obligations in accordance with the Contract;

AND WHEREAS we have agreed to give the Contractor a Guarantee:

THEREFORE WE hereby affirm that we are Guarantor and responsible to you, on behalf of the Contractor, up to a total of _____ (Amount of the guarantee in words and figures), and we undertake to pay you, upon your first written demand declaring the Contractor to be in default under the Contract, and without fail or argument, any sum or sums as specified by you, within the limits of _____ (Amount of Guarantee) as aforesaid without your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

This guarantee is valid until expiry of warranty period and satisfactory report of inspection committee.

[NAME OF GUARANTOR]

Name

Signature

Title

Address

Seal

Agreement No.-----

Dated: -----

CONDITIONS OF CONTRACT

1. Definitions

1.1 In this Contract, the following terms shall be interpreted as indicated:

- 1.1.1 "Commencement Date of the Contract" means the date of signing of the Contract between the Purchaser and the Contractor.
 - 1.1.2 "Contract" means the agreement entered into between the Purchaser and the Contractor, as recorded in the Contract Form signed by the parties, including all Schedules and Attachments thereto and all documents incorporated by reference therein.
 - 1.1.3 "Contract Price" means the price payable to the Contractor under the Contract for the full and proper performance of its contractual obligations.
 - 1.1.4 "Contractor" means the individual or firm / company / authorized dealer / supplier whose bid has been accepted by the Purchaser and the legal successors, in title to the Contractor.
 - 1.1.5 "Contract Value" means that portion of the Contract Price adjusted to give effect to such additions or deductions as are provided for in the Contract which is properly apportionable to the Goods or Services in question.
 - 1.1.6 "Defects Liability Expiry Certificate" means the certificate to be issued by the Client to the Contractor.
 - 1.1.7 "Defects Liability Period (Warranty Period)" means the period stated in the Schedule A Special Stipulations, following the taking over, during which the Contractor is responsible for making good defects and damage.
 - 1.1.8.1 "Client" means the specific Punjab Government Department for which the goods have been procured or any other person for the time being or from time to time duly appointed in writing by the Purchaser to act as Client for the purposes of the Contract.
 - 1.1.9 "Client's Representative" means any representative of the Client appointed from time to time by the Client.
 - 1.1.10 "Goods" means all of the equipment, and/or other materials which the Contractor is required to supply to the Purchaser under the Contract.
 - 1.1.11 "Purchaser" means the
Project Director,
Implementation of Master Plan of Safari Zoo Lahore,
Raiwind Road, Lahore.
 - 1.1.12 "Services" means services ancillary to the supply of the Goods, such as insurance, and any other incidental services, such as sexing, training and other such obligations of the Contractor covered under the Contract.
 - 1.1.13 "Taking-Over Certificate" means the certificate to be given by the Client to the Contractor.
 - 1.1.14 "Works" means all Service to be provided and work to be done by the Contractor under the Contract.
2. A bid security amounting **Rs. 5% of estimated cost** in favour of the Project Director,

Implementation of Master Plan of Safari Zoo Lahore,
Raiwind Road, Lahore

must accompany with tender in shape of Call Deposit Receipt (CDR) issued by any scheduled bank.

The bidder will purchase one set of tors documents for each lot and will clearly mention and tick mark the respective lot for which he is applying. Lot number shall also be mentioned on the envelope for technical and financial proposal.

3. Bid validity of offer should not be less than 90 days from date of opening of tenders.
4. **Warranty period must be mentioned clearly and it should not be less than 01 year after supply along with all required documents.**
14. The supplier will be bound to install the Equipments and will train the relevant staff of the safari zoo Lahore.
15. The selected bidders will be bound to provide 01 year back up service and 01 year maintenance of medical equipment's.
5. **List of previous work/experience shall also be provided with the tender.**
6. **Firms are required to quote only those items which strictly confirm to the specifications, otherwise the whole bid would be rejected.**
7. The rates should be inclusive of all applicable taxes.
8. Any erasing/cutting, overwriting, incomplete and ambiguous bids will not be considered.
9. **Successful Tenderers will be bound for providing all the Medical Equipments (Lot) within 90 days after issuance of work order, strictly according to the description / specifications mentioned in Tender. In case of any delay beyond the prescribed period, Project Director, Implementation of Master Plan of Safari Zoo Lahore, Raiwind Road, Lahore can extend the period as deem appropriate keeping in view the valid reason / justification.**
10. **In case of failure to complete the job within stipulated time/refused by the firm/rejection due to deviation from specifications/terms and conditions, the performance guarantee will be forfeited along with blacklisting of the firm.**
11. **There will be an inspection of medical equipment's and the required documents, at the completion of delivery.**
12. **Defective medical equipments can be rejected on the time of delivery by the authorized person or committee.**
13. **Successful Tenderers will be required to execute an agreement with the client (Project Director, Implementation of Master Plan of Safari Zoo Lahore, Raiwind Road, Lahore) on stamp paper according to the prescribed Government Rules within seven (07) days after approval of the bid as per settled terms & conditions and have to furnish the performance guarantee equal to the 10% in the form of Call Deposit Receipt (CDR) of the total tender amount, failing which the bid security will be forfeited.**
14. **Work order will be issued to the successful tenderer after approval of the Competent Authority.**
15. **No escalation etc. will be admissible after the issuance of work order in any case.**
16. **Payment will be made after completion of the work and satisfactory inspection report.**
17. **Income Tax/GST etc. will be deducted as per Government rules.**

18. Bids will be evaluated on the basis of criteria prescribed in the Bidding Documents.

19. In case of advance drawl from Government treasury / account, the contractor/ supplier should have to furnish a bank guarantee of equal amount from his own Bank Account which will be released after successful execution of the contract.

Certificate:

I/We do have read all above terms and conditions and accepted as such.

Name of Tenderer: _____

CNIC No.: _____

Address: _____

Contact No.: Office: _____ Cell: _____ NTN No.:

Date of Tender: _____ Signature: _____